



Peptide Test LLC

Product Information

Product Name	10ml Amber Sterile Vial	Specification	10ml,20mm
Batch No.	SV-A10C-AC240914	Mfg. Date	Sep.14 2024
Report Date	Sep.30 2024	Exp. Date	Sep.14 2027
Customer			

Certificate of Analysis

Items	Method	Specification	Result
Appearance	Visual inspection	Integrity, no obvious spots and scratches	Integrity, no obvious spots and scratches
Test for sterility	Current USP<71> Current EP<2.6.1>	Passed test	Passed test
Bacterial Endotoxin	Current USP<85> Current EP<2.6.14>	< 0.03 IU/Vial	Conforms
Particles, $\geq 25 \mu\text{m}$	Current USP<788> Current EP<2.9.19>	< 150	0
Particles, $\geq 10 \mu\text{m}$	Current USP<788> Current EP<2.9.19>	< 1500	5
Sealing Capacity	In house	No leak	No leak