



# Peptide Test LLC

## Product Information

|              |                        |               |            |
|--------------|------------------------|---------------|------------|
| Product Name | 2ml Amber Sterile vial | Specification | 2ml, 13mm  |
| Batch No.    | SV-A2-AC241209         | Mfg. Date     | Dec.9 2024 |
| Report Date  | Jan.3 2025             | Exp. Date     | Dec.9 2027 |
| Customer     |                        |               |            |

## Certificate of Analysis

| Items               | Method                                 | Specification                             | Result                                    |
|---------------------|----------------------------------------|-------------------------------------------|-------------------------------------------|
| Appearance          | Visual inspection                      | Integrity, no obvious spots and scratches | Integrity, no obvious spots and scratches |
| Test for sterility  | Current USP<71><br>Current EP<2.6.1>   | Passed test                               | Passed test                               |
| Bacterial Endotoxin | Current USP<85><br>Current EP<2.6.14>  | < 0.03 IU/Vial                            | Conforms                                  |
| Particles, ≥25 µm   | Current USP<788><br>Current EP<2.9.19> | < 150                                     | 0                                         |
| Particles, ≥10 µm   | Current USP<788><br>Current EP<2.9.19> | < 1500                                    | 1                                         |
| Sealing Capacity    | In house                               | No leak                                   | Conforms                                  |